
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008169
9:00 16/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:38
STERILIZING 4:40
136.4 C 220kPa 4:40
135.9 C 220kPa 5:10
136.0 C 220kPa 5:40
136.3 C 223kPa 6:10
136.1 C 221kPa 6:40
136.2 C 219kPa 7:10
136.4 C 223kPa 7:40
136.3 C 220kPa 8:10
136.3 C 220kPa 8:10
VENTING 8:11
AIR DRYING 9:04
CYCLE COMPLETE 25:58

Digital Signature #
3D496673B9EC2E3B
