
Midmark M11 - v1.0.5
Total Cycles: 3331

Name: RITTER2
Sterilizer ID: V2313228
01/29/2024 12:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 159.1 F 2.6
4:00 175.6 F 6.4
6:00 199.0 F 11.3
8:00 222.8 F 6.1
10:00 236.1 F 11.4
12:00 235.0 F 8.9
14:00 238.4 F 9.9
16:00 248.2 F 14.7
18:00 261.0 F 21.8
20:00 272.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.4 F 29.5
2:00 271.2 F 28.7
3:00 272.4 F 29.6
4:00 272.1 F 29.1

Min 271.1 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 238.6 F 0.0
10:00 203.2 F 0.0
15:00 191.4 F 0.0
20:00 187.6 F 0.0
25:00 185.4 F 0.0
30:00 185.0 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

01/29/2024 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____