
Midmark M11 - v1.0.0
Total Cycles: 4925

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2024 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.3 F 0.0
2:00 153.1 F 0.1
4:00 170.1 F 0.4
6:00 193.3 F 2.1
8:00 209.4 F 5.5
10:00 218.8 F 9.2
12:00 231.3 F 13.1
14:00 232.5 F 8.2
16:00 242.9 F 12.8
18:00 239.9 F 10.4
20:00 237.5 F 8.6
22:00 249.2 F 14.5
24:00 259.9 F 20.5
26:00 269.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 271.7 F 28.3
2:00 272.3 F 28.7
3:00 272.5 F 28.8
4:00 272.4 F 28.6

Min 271.2 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 185.3 F 0.0
10:00 191.0 F 0.0
15:00 183.1 F 0.0
20:00 182.6 F 0.0
25:00 187.2 F 0.0
30:00 196.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 27:16
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:05:53
CYCLE COMPLETE

09/20/2024 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____