
Midmark M11 - v1.0.5
Total Cycles: 4186

Name: RITTER2
Sterilizer ID: V2313228
12/04/2024 06:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.8 F 0.0
2:00 175.2 F 1.9
4:00 201.7 F 7.5
6:00 219.7 F 13.3
8:00 230.9 F 9.2
10:00 235.5 F 9.3
12:00 243.7 F 12.3
14:00 245.7 F 13.6
16:00 259.2 F 20.9
18:00 271.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.4 F 29.2
2:00 272.9 F 29.6
3:00 271.9 F 29.3
4:00 272.3 F 29.2

Min 271.7 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 231.9 F 0.0
10:00 203.4 F 0.0
15:00 189.9 F 0.0
20:00 185.0 F 0.0
25:00 183.5 F 0.0
30:00 183.1 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 18:38
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

12/04/2024 07:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____