

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1733

Name: RITTER 1  
Sterilizer ID: V2178198  
03/23/2022 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.4 F 0.0  
2:00 129.3 F 0.0  
4:00 140.5 F 1.3  
6:00 167.8 F 3.9  
8:00 200.3 F 7.4  
10:00 212.6 F 12.5  
12:00 227.0 F 6.7  
14:00 236.3 F 11.8  
16:00 235.9 F 8.7  
18:00 240.4 F 10.1  
20:00 246.1 F 13.0  
22:00 258.8 F 19.9  
24:00 270.3 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.6  
1:00 272.2 F 28.5  
2:00 272.6 F 28.7  
3:00 272.6 F 28.9  
4:00 272.8 F 29.2

Min 271.9 F 28.5  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.6 F 0.0  
5:00 234.8 F 0.0  
10:00 218.1 F 0.0  
15:00 214.0 F 0.0  
20:00 227.4 F 0.0  
25:00 231.0 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:28  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:03:28  
CYCLE COMPLETE

03/23/2022 10:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_