
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008272
10:14 30/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:38
STERILIZING 3:46
136.4 C 220kPa 3:46
136.2 C 223kPa 4:16
136.0 C 222kPa 4:46
136.2 C 222kPa 5:16
136.4 C 223kPa 5:46
136.4 C 222kPa 6:16
136.4 C 222kPa 6:46
136.3 C 222kPa 7:16
136.3 C 222kPa 7:16
VENTING 7:16
CYCLE ABORTED 8:06
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:06

Digital Signature #
5B5E23BB3279FC21
