
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006694
15:35 05/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:51
STERILIZING 3:55
136.4 C 220kPa 3:55
136.3 C 224kPa 4:25
136.4 C 223kPa 4:55
136.4 C 225kPa 5:25
136.4 C 223kPa 5:55
136.4 C 223kPa 6:25
136.3 C 222kPa 6:55
136.4 C 224kPa 7:25
136.4 C 224kPa 7:25
VENTING 7:26
AIR DRYING 8:14
CYCLE COMPLETE 8:44

Digital Signature #
12273C02F0ACE898
