
Midmark M11 - v1.0.0
Total Cycles: 2962

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2023 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.5 F 0.0
2:00 180.7 F 4.0
4:00 194.5 F 8.5
6:00 213.6 F 13.5
8:00 223.7 F 8.3
10:00 237.1 F 12.5
12:00 238.8 F 10.8
14:00 237.3 F 9.1
16:00 251.5 F 16.1
18:00 263.4 F 23.1
20:00 271.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.2
2:00 271.9 F 28.5
3:00 272.0 F 29.0
4:00 272.6 F 29.1

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 222.4 F 0.0
10:00 239.8 F 0.0
15:00 232.5 F 0.0
20:00 231.5 F 0.0
25:00 226.8 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

03/01/2023 01:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____