
Midmark M11 - v1.0.0
Total Cycles: 3201

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2023 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 181.0 F 2.5
4:00 191.8 F 6.7
6:00 211.1 F 12.1
8:00 224.0 F 7.2
10:00 237.4 F 12.6
12:00 237.5 F 10.0
14:00 236.1 F 8.4
16:00 250.2 F 15.4
18:00 262.5 F 22.5
20:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.4 F 29.1
2:00 272.1 F 28.6
3:00 271.7 F 28.7
4:00 272.6 F 29.0

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 209.8 F 0.0
10:00 222.5 F 0.0
15:00 228.9 F 0.0
20:00 228.5 F 0.0
25:00 222.3 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

05/04/2023 10:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____