
Midmark M11 - v1.0.0
Total Cycles: 1553

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2022 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.4 F 0.0
2:00 187.0 F 4.3
4:00 191.6 F 8.9
6:00 216.2 F 7.6
8:00 225.9 F 9.3
10:00 236.4 F 9.1
12:00 240.7 F 12.2
14:00 239.9 F 10.7
16:00 253.2 F 17.2
18:00 264.7 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.8 F 29.5
2:00 272.1 F 28.9
3:00 271.7 F 28.9
4:00 272.8 F 29.5

Min 271.5 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 230.8 F 0.0
10:00 238.1 F 0.0
15:00 231.9 F 0.0
20:00 226.2 F 0.0
25:00 227.1 F 0.0
30:00 227.6 F 0.0

Drying Complete

FILLING: 1:20
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

01/24/2022 11:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____