
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007758
7:42 19/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:02
STERILIZING 4:08
136.5 C 222kPa 4:08
136.3 C 222kPa 4:38
136.1 C 220kPa 5:08
136.2 C 221kPa 5:38
136.3 C 220kPa 6:08
136.4 C 223kPa 6:38
136.4 C 220kPa 7:08
136.0 C 220kPa 7:38
136.0 C 220kPa 7:38
VENTING 7:38
AIR DRYING 8:28
CYCLE COMPLETE 17:02

Digital Signature #
78D571717D74766C
