
Midmark M11 - v1.0.0
Total Cycles: 4656

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2024 08:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 71.3 F 0.0
4:00 86.0 F 0.1
6:00 113.0 F 0.3
8:00 143.8 F 1.1
10:00 170.0 F 4.1
12:00 192.8 F 7.3
14:00 217.8 F 10.6
16:00 213.4 F 7.9
18:00 224.7 F 6.8
20:00 234.0 F 10.6
22:00 233.1 F 8.3
24:00 241.7 F 11.3
26:00 236.7 F 8.8
28:00 249.7 F 14.9
30:00 260.3 F 20.7
32:00 270.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.4
2:00 272.2 F 28.7
3:00 271.5 F 28.0
4:00 272.3 F 28.7

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 219.6 F 0.0
10:00 191.1 F 0.0
15:00 190.3 F 0.0
20:00 187.5 F 0.0
25:00 186.4 F 0.0
30:00 186.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 32:57
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:11:13
CYCLE COMPLETE

06/19/2024 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____