

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3523

Name: RITTER2  
Sterilizer ID: V2313228  
04/12/2024 02:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 107.7 F 0.0  
2:00 110.5 F 0.1  
4:00 132.2 F 0.5  
6:00 172.0 F 3.4  
8:00 199.5 F 7.4  
10:00 209.8 F 11.3  
12:00 220.4 F 4.8  
14:00 227.0 F 9.1  
16:00 238.0 F 12.2  
18:00 239.9 F 11.2  
20:00 237.3 F 9.8  
22:00 250.5 F 16.1  
24:00 261.5 F 22.5  
26:00 271.8 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.7  
1:00 271.9 F 29.5  
2:00 272.4 F 29.7  
3:00 272.7 F 30.0  
4:00 272.4 F 30.2

Min 271.6 F 29.5  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.2 F 0.0  
5:00 230.2 F 0.0  
10:00 212.5 F 0.0  
15:00 211.3 F 0.0  
20:00 204.3 F 0.0  
25:00 195.8 F 0.0  
30:00 196.1 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:16  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:05:51  
CYCLE COMPLETE

04/12/2024 03:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_