

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002520  
15:01 25/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:17  
PRESSURIZING 1:22  
STERILIZING 2:22  
136.0 C 220kPa 2:22  
136.2 C 221kPa 2:52  
136.4 C 223kPa 3:22  
136.3 C 225kPa 3:52  
136.2 C 222kPa 4:22  
136.4 C 224kPa 4:52  
136.1 C 222kPa 5:22  
136.4 C 224kPa 5:52  
136.4 C 224kPa 5:52  
VENTING 5:52  
AIR DRYING 6:47  
CYCLE COMPLETE 6:48

-----  
Digital Signature #  
1EEE8644D8896933  
-----