
Midmark M11 - v1.0.0
Total Cycles: 5059

Name: RITTER 1
Sterilizer ID: V2178198
11/05/2024 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.0 F 0.0
2:00 120.1 F 0.1
4:00 149.4 F 0.5
6:00 179.6 F 1.6
8:00 200.7 F 5.4
10:00 221.6 F 9.2
12:00 233.4 F 13.2
14:00 233.2 F 8.4
16:00 240.7 F 13.1
18:00 239.2 F 10.4
20:00 235.9 F 8.3
22:00 249.1 F 14.7
24:00 260.5 F 21.0
26:00 271.0 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.4 F 28.9
2:00 271.7 F 28.3
3:00 271.8 F 28.8
4:00 272.5 F 29.0

Min 270.9 F 28.0
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.3 F 0.0
5:00 235.5 F 0.0
10:00 191.7 F 0.0
15:00 190.8 F 0.0
20:00 190.4 F 0.0
25:00 189.7 F 0.0
30:00 190.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:47
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:05:23
CYCLE COMPLETE

11/05/2024 04:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____