

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1137

Name: RITTER2  
Sterilizer ID: V2313228  
05/30/2022 11:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.5 F 0.0  
2:00 184.8 F 3.8  
4:00 193.3 F 8.7  
6:00 209.9 F 9.7  
8:00 226.3 F 9.2  
10:00 233.7 F 9.6  
12:00 241.7 F 12.8  
14:00 242.2 F 12.0  
16:00 255.5 F 18.9  
18:00 267.0 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.6 F 29.8  
2:00 271.5 F 29.0  
3:00 272.2 F 29.9  
4:00 272.4 F 29.6  
  
Min 271.2 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.5 F 0.0  
5:00 212.9 F 0.0  
10:00 233.1 F 0.0  
15:00 233.5 F 0.0  
20:00 228.4 F 0.0  
25:00 225.4 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 19:28  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

05/30/2022 12:06 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_