
Midmark M11 - v1.0.0
Total Cycles: 1141

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2021 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.3 F 0.0
2:00 139.3 F 1.0
4:00 163.5 F 4.0
6:00 189.4 F 8.6
8:00 212.5 F 10.5
10:00 222.0 F 7.1
12:00 234.7 F 12.0
14:00 235.3 F 8.7
16:00 240.0 F 10.4
18:00 245.7 F 13.1
20:00 258.1 F 19.7
22:00 269.0 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 273.0 F 29.2
2:00 271.6 F 28.4
3:00 272.7 F 29.3
4:00 272.6 F 28.9

Min 271.4 F 28.3
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 218.3 F 0.0
10:00 204.2 F 0.0
15:00 199.1 F 0.0
20:00 191.9 F 0.0
25:00 190.4 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 23:06
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:02:01
CYCLE COMPLETE

09/02/2021 12:51 PM
ACCEPTED: X
REJECTED:
Operator: U02