
Midmark M11 - v1.0.0
Total Cycles: 2705

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2022 06:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.3 F 0.0
2:00 198.7 F 4.0
4:00 204.8 F 9.0
6:00 222.5 F 6.5
8:00 234.2 F 10.5
10:00 237.2 F 7.4
12:00 242.2 F 11.1
14:00 246.6 F 13.5
16:00 259.6 F 20.6
18:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.4 F 29.1
2:00 271.9 F 28.4
3:00 272.0 F 28.8
4:00 272.4 F 28.8

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.5 F 0.0
5:00 218.3 F 0.0
10:00 235.0 F 0.0
15:00 236.3 F 0.0
20:00 228.7 F 0.0
25:00 222.6 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

12/20/2022 07:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____