
Midmark M11 - v1.0.5
Total Cycles: 2529

Name: RITTER2
Sterilizer ID: V2313228
06/19/2023 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.7 F 0.0
2:00 193.3 F 3.0
4:00 199.6 F 7.9
6:00 217.1 F 13.5
8:00 228.5 F 9.0
10:00 236.5 F 10.2
12:00 242.8 F 12.6
14:00 242.2 F 11.6
16:00 255.1 F 18.4
18:00 266.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.6 F 29.6
2:00 271.6 F 28.9
3:00 272.5 F 29.4
4:00 271.6 F 29.1

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 223.7 F 0.0
10:00 232.1 F 0.0
15:00 228.1 F 0.0
20:00 227.2 F 0.0
25:00 227.7 F 0.0
30:00 224.6 F 0.0

Drying Complete

FILLING: 1:57
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:38
CYCLE COMPLETE

06/19/2023 11:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____