
Midmark M11 - v1.0.5
Total Cycles: 3053

Name: RITTER2
Sterilizer ID: V2313228
10/25/2023 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.0 F 0.0
2:00 151.0 F 2.3
4:00 177.0 F 6.0
6:00 197.8 F 10.4
8:00 211.2 F 5.2
10:00 224.5 F 8.3
12:00 237.3 F 13.3
14:00 238.7 F 11.0
16:00 238.0 F 9.8
18:00 252.1 F 16.6
20:00 264.0 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.2 F 29.1
2:00 272.2 F 29.4
3:00 272.0 F 28.8
4:00 272.5 F 29.5

Min 271.4 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.4 F 0.0
5:00 195.9 F 0.0
10:00 194.6 F 0.0
15:00 193.3 F 0.0
20:00 191.2 F 0.0
25:00 189.6 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

10/25/2023 12:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____