

-----  
Midmark M11 - v1.0.5  
Total Cycles: 106

Name: RITTER2  
Sterilizer ID: V2313228  
06/16/2021 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.2 F 0.0  
2:00 182.5 F 2.3  
4:00 191.1 F 5.8  
6:00 206.2 F 10.2  
8:00 224.7 F 6.4  
10:00 232.3 F 9.5  
12:00 237.1 F 10.2  
14:00 242.4 F 12.2  
16:00 240.6 F 11.1  
18:00 253.0 F 17.4  
20:00 264.0 F 23.8  
22:00 271.9 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.4 F 29.4  
2:00 271.1 F 28.9  
3:00 272.8 F 30.0  
4:00 272.2 F 29.4

Min 271.0 F 28.8  
Max 273.2 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 227.3 F 0.0  
10:00 228.9 F 0.0  
15:00 229.5 F 0.0  
20:00 228.5 F 0.0  
25:00 236.5 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 22:08  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:01:06  
CYCLE COMPLETE

06/16/2021 10:37 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_