
Midmark M11 - v1.0.0
Total Cycles: 1229

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2021 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.2 F 0.0
2:00 163.8 F 2.9
4:00 190.6 F 7.4
6:00 211.7 F 12.7
8:00 220.2 F 8.0
10:00 233.8 F 13.1
12:00 238.5 F 11.0
14:00 237.4 F 9.1
16:00 251.3 F 15.8
18:00 263.2 F 22.6
20:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.4 F 29.4
2:00 272.5 F 29.8
3:00 272.1 F 29.1
4:00 272.9 F 29.8

Min 271.8 F 29.0
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 241.4 F 0.0
10:00 228.3 F 0.0
15:00 228.5 F 0.0
20:00 227.8 F 0.0
25:00 229.9 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

10/05/2021 06:29 AM
ACCEPTED: X
REJECTED:
Operator: U01