

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3688

Name: RITTER2  
Sterilizer ID: V2313228  
06/12/2024 10:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.0 F 0.0  
2:00 137.6 F 1.3  
4:00 141.8 F 3.8  
6:00 164.1 F 7.2  
8:00 185.3 F 11.2  
10:00 218.8 F 4.7  
12:00 225.8 F 8.0  
14:00 234.0 F 12.5  
16:00 233.1 F 9.1  
18:00 241.2 F 11.0  
20:00 244.1 F 12.9  
22:00 255.3 F 18.8  
24:00 265.4 F 25.0  
26:00 271.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.7  
1:00 272.8 F 29.9  
2:00 272.3 F 30.3  
3:00 271.7 F 29.7  
4:00 272.3 F 30.1

Min 271.6 F 29.6  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 235.2 F 0.0  
10:00 190.5 F 0.0  
15:00 183.9 F 0.0  
20:00 178.8 F 0.0  
25:00 178.2 F 0.0  
30:00 177.7 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 26:41  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:05:13  
CYCLE COMPLETE

06/12/2024 11:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_