
Midmark M11 - v1.0.5
Total Cycles: 668

Name: RITTER2
Sterilizer ID: V2313228
12/30/2021 07:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 76.4 F 0.0
4:00 80.1 F 0.1
6:00 108.3 F 0.2
8:00 136.8 F 2.5
10:00 162.9 F 5.3
12:00 186.3 F 8.7
14:00 204.1 F 12.5
16:00 221.1 F 6.1
18:00 230.0 F 10.6
20:00 233.8 F 7.9
22:00 242.5 F 13.0
24:00 241.7 F 11.6
26:00 254.0 F 17.8
28:00 264.8 F 24.3
30:00 271.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.4
1:00 272.0 F 29.5
2:00 271.9 F 29.7
3:00 272.1 F 29.8
4:00 272.1 F 29.8

Min 271.8 F 29.3
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 220.2 F 0.0
10:00 236.0 F 0.0
15:00 227.5 F 0.0
20:00 221.9 F 0.0
25:00 220.9 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:28
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

12/30/2021 09:18 AM
ACCEPTED: X
REJECTED:
Operator: U02