
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005460
10:19 02/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:33
STERILIZING 3:38
136.4 C 222kPa 3:38
136.3 C 224kPa 4:08
136.4 C 224kPa 4:38
136.4 C 223kPa 5:08
136.2 C 224kPa 5:38
136.3 C 222kPa 6:08
136.3 C 222kPa 6:38
136.0 C 221kPa 7:08
136.0 C 221kPa 7:08
VENTING 7:08
AIR DRYING 8:03
CYCLE COMPLETE 8:10

Digital Signature #
84C9CE3A38FCE3B9
