
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005614
10:31 29/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:01
STERILIZING 3:02
136.5 C 223kPa 3:02
136.2 C 224kPa 3:32
136.2 C 224kPa 4:02
136.2 C 222kPa 4:32
136.2 C 222kPa 5:02
136.4 C 223kPa 5:32
136.4 C 224kPa 6:02
136.4 C 224kPa 6:32
136.4 C 224kPa 6:32
VENTING 6:33
AIR DRYING 7:23
CYCLE COMPLETE 7:34

Digital Signature #
881C8030FEE40EEA
