
Midmark M11 - v1.0.0
Total Cycles: 2517

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2022 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.4 F 0.0
2:00 178.4 F 2.7
4:00 176.7 F 6.5
6:00 196.5 F 11.2
8:00 216.0 F 5.2
10:00 229.1 F 10.3
12:00 231.5 F 7.5
14:00 243.3 F 13.0
16:00 241.8 F 11.4
18:00 254.8 F 17.9
20:00 266.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.2 F 28.9
2:00 272.1 F 29.1
3:00 271.8 F 28.6
4:00 272.4 F 29.3

Min 271.5 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 217.4 F 0.0
5:00 232.6 F 0.0
10:00 237.8 F 0.0
15:00 232.3 F 0.0
20:00 219.7 F 0.0
25:00 216.8 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

11/07/2022 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____