
Midmark M11 - v1.0.0
Total Cycles: 2400

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2022 05:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 173.9 F 3.1
4:00 190.2 F 6.8
6:00 219.4 F 11.8

C383: HEATUP MODE
DOOR OPEN
CHAMBER TEMP 220F
CHAMBER PRESS 11.9PSI
** ITEMS NOT STERILE **

10/04/2022 05:07 PM