
Midmark M11 - v1.0.5
Total Cycles: 2677

Name: RITTER2
Sterilizer ID: V2313228
07/20/2023 02:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.2 F 0.0
2:00 152.8 F 0.9
4:00 164.8 F 4.3
6:00 190.2 F 8.8
8:00 208.4 F 9.4
10:00 218.8 F 9.2
12:00 233.0 F 9.9
14:00 237.9 F 12.6
16:00 240.5 F 11.3
18:00 252.8 F 17.8
20:00 263.5 F 24.7
22:00 270.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.3
1:00 271.9 F 30.4
2:00 272.0 F 30.5
3:00 271.8 F 30.5
4:00 271.8 F 30.4

Min 271.6 F 30.1
Max 272.9 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 232.2 F 0.0
10:00 228.8 F 0.0
15:00 233.7 F 0.0
20:00 223.2 F 0.0
25:00 225.2 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:03
CYCLE COMPLETE

07/20/2023 03:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____