
Midmark M11 - v1.0.0
Total Cycles: 4962

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2024 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.7 F 0.0
2:00 126.9 F 0.4
4:00 147.4 F 0.8
6:00 176.8 F 3.7
8:00 211.2 F 7.7
10:00 217.4 F 12.1
12:00 225.4 F 6.0
14:00 232.2 F 10.8
16:00 232.2 F 7.2
18:00 242.7 F 13.1
20:00 242.3 F 11.4
22:00 254.6 F 17.7
24:00 265.5 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.2 F 29.0
2:00 272.3 F 28.8
3:00 272.3 F 29.2
4:00 272.4 F 28.9

Min 270.9 F 28.1
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 242.5 F 0.0
10:00 225.2 F 0.0
15:00 215.0 F 0.0
20:00 227.1 F 0.0
25:00 233.5 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:54
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

10/03/2024 05:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____