

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008066  
10:23 14/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:20  
PRESSURIZING 2:22  
STERILIZING 3:24  
136.4 C 220kPa 3:24  
136.4 C 225kPa 3:54  
136.4 C 223kPa 4:24  
136.5 C 225kPa 4:54  
136.2 C 221kPa 5:24  
136.3 C 224kPa 5:54  
136.2 C 224kPa 6:24  
136.3 C 224kPa 6:54  
136.3 C 224kPa 6:54  
VENTING 6:54  
AIR DRYING 7:45  
CYCLE COMPLETE 37:45

-----  
Digital Signature #  
2FD21E35BE9F33A0  
-----