

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001360  
9:13 22/03/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:40  
PRESSURIZING 3:41  
STERILIZING 4:57  
136.5 C 220kPa 4:57  
136.3 C 224kPa 5:27  
136.4 C 224kPa 5:57  
136.3 C 223kPa 6:27  
136.4 C 224kPa 6:57  
136.2 C 222kPa 7:27  
136.2 C 222kPa 7:57  
136.1 C 223kPa 8:27  
136.1 C 223kPa 8:27  
VENTING 8:27  
AIR DRYING 9:21  
CYCLE COMPLETE 9:22

-----  
Digital Signature #  
CE8B0FDC4E0C6810  
-----