
Midmark M11 - v1.0.0
Total Cycles: 4503

Name: RITTER 1
Sterilizer ID: V2178198
04/29/2024 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.7 F 0.0
2:00 105.3 F 0.9
4:00 123.5 F 3.5
6:00 154.8 F 6.8
8:00 183.0 F 10.8
10:00 207.4 F 3.8
12:00 217.6 F 7.6
14:00 231.2 F 12.1
16:00 230.7 F 8.0
18:00 243.0 F 13.4
20:00 241.5 F 11.3
22:00 253.8 F 17.5
24:00 264.8 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.0
2:00 271.4 F 28.7
3:00 272.3 F 29.0
4:00 271.4 F 28.6

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 213.2 F 0.0
10:00 213.7 F 0.0
15:00 228.1 F 0.0
20:00 228.8 F 0.0
25:00 235.5 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:55
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:04:13
CYCLE COMPLETE

04/29/2024 04:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____