

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4994

Name: RITTER 1  
Sterilizer ID: V2178198  
10/15/2024 08:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.6 F 0.0  
2:00 75.0 F 0.0  
4:00 114.1 F 0.1  
6:00 117.5 F 0.1  
8:00 150.1 F 0.3  
10:00 176.5 F 0.9  
12:00 197.9 F 4.3  
14:00 209.9 F 7.4  
16:00 222.0 F 10.9  
18:00 225.5 F 6.5  
20:00 235.0 F 9.8  
22:00 238.8 F 9.2  
24:00 243.4 F 12.3  
26:00 241.1 F 10.8  
28:00 254.1 F 17.3  
30:00 265.3 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.1  
1:00 272.5 F 28.9  
2:00 271.7 F 28.0  
3:00 272.7 F 28.9  
4:00 272.0 F 28.3  
  
Min 271.5 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 213.8 F 0.0  
10:00 212.9 F 0.0  
15:00 226.9 F 0.0  
20:00 232.1 F 0.0  
25:00 223.4 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 0:43  
HEATING: 31:39  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:09:48  
CYCLE COMPLETE

10/15/2024 09:37 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_