

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005559  
8:39 17/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:28  
STERILIZING 3:14  
136.0 C 220kPa 3:14  
136.2 C 220kPa 3:45  
136.2 C 221kPa 4:15  
136.4 C 224kPa 4:45  
136.3 C 223kPa 5:15  
136.1 C 220kPa 5:45  
136.4 C 224kPa 6:15  
136.4 C 223kPa 6:45  
136.4 C 223kPa 6:45  
VENTING 6:45  
AIR DRYING 7:31  
CYCLE COMPLETE 10:52

-----  
Digital Signature #  
43256EEA227A90B5  
-----