
Midmark M11 - v1.0.5
Total Cycles: 2255

Name: RITTER2
Sterilizer ID: V2313228
04/14/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.1 F 0.0
2:00 118.2 F 0.0
4:00 137.8 F 0.4
6:00 169.7 F 3.7
8:00 197.9 F 7.4
10:00 214.0 F 11.8
12:00 221.1 F 6.1
14:00 229.6 F 10.7
16:00 231.9 F 7.7
18:00 242.4 F 12.7
20:00 240.8 F 11.3
22:00 253.7 F 17.6
24:00 264.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.5 F 29.3
2:00 271.3 F 29.1
3:00 272.3 F 29.3
4:00 271.2 F 28.8

Min 271.1 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 219.6 F 0.0
10:00 197.6 F 0.0
15:00 192.9 F 0.0
20:00 190.9 F 0.0
25:00 188.5 F 0.0
30:00 189.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:56
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:04:15
CYCLE COMPLETE

04/14/2023 02:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____