

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3682

Name: RITTER2  
Sterilizer ID: V2313228  
06/10/2024 02:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 96.9 F 0.0  
2:00 98.2 F 0.4  
4:00 110.7 F 1.9  
6:00 136.5 F 4.4  
8:00 163.2 F 7.5  
10:00 191.2 F 11.1  
12:00 207.5 F 4.4  
14:00 213.1 F 6.6  
16:00 223.2 F 10.3  
18:00 231.1 F 10.1  
20:00 236.3 F 10.9  
22:00 234.6 F 8.4  
24:00 246.6 F 14.3  
26:00 257.1 F 20.1  
28:00 267.0 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.3 F 30.1  
2:00 272.3 F 29.8  
3:00 271.4 F 29.3  
4:00 271.2 F 29.5

Min 271.1 F 29.2  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.2 F 0.0  
5:00 199.1 F 0.0  
10:00 195.1 F 0.0  
15:00 189.9 F 0.0  
20:00 193.7 F 0.0  
25:00 196.2 F 0.0  
30:00 196.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 29:39  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:08:19  
CYCLE COMPLETE

06/10/2024 03:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_