
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003196
10:16 23/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:26
STERILIZING 3:25
136.4 C 221kPa 3:25
136.4 C 223kPa 3:55
136.4 C 219kPa 4:25
136.2 C 223kPa 4:55
136.4 C 223kPa 5:25
135.7 C 219kPa 5:55
136.4 C 223kPa 6:25
136.4 C 221kPa 6:55
136.4 C 221kPa 6:55
VENTING 6:55
AIR DRYING 7:49
CYCLE COMPLETE 7:59

Digital Signature #
614A56379924382C
