
Midmark M11 - v1.0.0
Total Cycles: 2045

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2022 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.1 F 0.0
2:00 180.6 F 3.3
4:00 188.5 F 7.2
6:00 208.8 F 12.2
8:00 224.1 F 6.8
10:00 237.8 F 11.8
12:00 234.2 F 8.6
14:00 240.8 F 10.3
16:00 246.2 F 13.4
18:00 258.6 F 20.0
20:00 269.6 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.1
2:00 271.9 F 28.5
3:00 271.5 F 28.5
4:00 272.6 F 29.2

Min 271.2 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 234.2 F 0.0
10:00 235.2 F 0.0
15:00 229.3 F 0.0
20:00 219.6 F 0.0
25:00 217.1 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

06/22/2022 12:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____