

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2954

Name: RITTER 1  
Sterilizer ID: V2178198  
02/28/2023 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.5 F 0.0  
2:00 122.4 F 0.0  
4:00 148.5 F 0.5  
6:00 179.5 F 3.6  
8:00 203.4 F 7.2  
10:00 217.2 F 11.8  
12:00 224.1 F 6.7  
14:00 236.4 F 11.8  
16:00 235.2 F 9.0  
18:00 239.3 F 9.7  
20:00 247.5 F 14.1  
22:00 260.0 F 21.0  
24:00 271.3 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.2  
2:00 271.5 F 28.4  
3:00 272.6 F 29.3  
4:00 272.1 F 28.7

Min 271.4 F 28.3  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 231.5 F 0.0  
10:00 231.2 F 0.0  
15:00 222.3 F 0.0  
20:00 229.7 F 0.0  
25:00 233.3 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:41  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:02:47  
CYCLE COMPLETE

02/28/2023 02:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_