
Midmark M11 - v1.0.5
Total Cycles: 1459

Name: RITTER2
Sterilizer ID: V2313228
09/06/2022 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 173.7 F 2.6
4:00 191.7 F 7.3
6:00 223.0 F 12.9
8:00 228.3 F 8.1
10:00 241.9 F 13.5
12:00 240.1 F 11.3
14:00 238.5 F 10.0
16:00 252.0 F 16.7
18:00 263.2 F 23.4
20:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.9 F 29.7
2:00 272.3 F 29.0
3:00 271.7 F 29.2
4:00 272.9 F 29.6

Min 271.3 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 231.9 F 0.0
10:00 227.4 F 0.0
15:00 233.2 F 0.0
20:00 232.7 F 0.0
25:00 229.7 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

09/06/2022 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____