
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000514
8:40 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:48
STERILIZING 5:04
136.2 C 220kPa 5:04
136.1 C 222kPa 5:34
136.3 C 224kPa 6:04
136.4 C 224kPa 6:34
136.2 C 222kPa 7:04
136.4 C 222kPa 7:34
136.4 C 222kPa 8:04
136.3 C 223kPa 8:34
136.3 C 223kPa 8:34
VENTING 8:34
AIR DRYING 9:30
CYCLE COMPLETE 9:49

Digital Signature #
995F12EAB9624056
