

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2370

Name: RITTER 1  
Sterilizer ID: V2178198  
09/26/2022 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
2:00 202.5 F 4.5  
4:00 208.1 F 9.6  
6:00 220.9 F 6.7  
8:00 232.2 F 10.6  
10:00 235.8 F 7.6  
12:00 241.7 F 11.4  
14:00 245.4 F 13.2  
16:00 258.3 F 20.2  
18:00 270.0 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.6  
1:00 272.4 F 29.3  
2:00 271.5 F 28.6  
3:00 271.6 F 29.1  
4:00 272.3 F 29.3  
  
Min 271.1 F 28.5  
Max 272.4 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.2 F 0.0  
5:00 236.6 F 0.0  
10:00 231.8 F 0.0  
15:00 229.7 F 0.0  
20:00 228.2 F 0.0  
25:00 229.6 F 0.0  
30:00 229.7 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 18:55  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:12  
CYCLE COMPLETE

09/26/2022 02:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_