
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000908
15:57 04/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:22
STERILIZING 4:33
136.3 C 231kPa 4:33
136.4 C 233kPa 5:03
136.2 C 230kPa 5:33
136.4 C 232kPa 6:03
136.4 C 233kPa 6:33
136.3 C 231kPa 7:03
136.1 C 230kPa 7:33
136.4 C 233kPa 8:03
136.4 C 233kPa 8:03
VENTING 8:03
AIR DRYING 8:59
CASSETTE REMOVED
CYCLE COMPLETE 9:04

Digital Signature #
8611CD663D4098E2
