
Midmark M11 - v1.0.5
Total Cycles: 1126

Name: RITTER2
Sterilizer ID: V2313228
05/27/2022 06:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 74.4 F 0.0
4:00 85.5 F 0.0
6:00 99.2 F 0.0
8:00 158.4 F 0.2
10:00 173.3 F 3.4
12:00 194.5 F 7.0
14:00 214.4 F 10.9
16:00 221.8 F 6.1
18:00 229.4 F 8.8
20:00 239.8 F 13.1
22:00 237.1 F 10.2
24:00 237.0 F 8.8
26:00 247.9 F 14.7
28:00 259.0 F 20.8
30:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.4 F 29.5
2:00 271.5 F 29.4
3:00 272.1 F 29.3
4:00 271.7 F 29.6

Min 271.2 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 198.8 F 0.0
10:00 211.0 F 0.0
15:00 221.6 F 0.0
20:00 200.2 F 0.0
25:00 204.5 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:07
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:59
CYCLE COMPLETE

05/27/2022 07:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____