

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 001233  
13:37 29/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:26  
PRESSURIZING 4:27  
STERILIZING 5:42  
136.0 C 222kPa 5:42  
136.4 C 225kPa 6:12  
136.1 C 222kPa 6:42  
136.4 C 225kPa 7:12  
136.4 C 223kPa 7:42  
136.4 C 224kPa 8:12  
136.4 C 224kPa 8:42  
136.2 C 222kPa 9:12  
136.2 C 222kPa 9:12  
VENTING 9:12  
AIR DRYING 10:09  
CYCLE COMPLETE 22:34  
-----

Digital Signature #  
7ADE591CD3989AC2  
-----