

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004471  
15:12 03/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:02  
STERILIZING 4:05  
136.4 C 222kPa 4:05  
136.2 C 219kPa 4:35  
136.4 C 225kPa 5:05  
136.3 C 225kPa 5:35  
136.2 C 224kPa 6:05  
136.4 C 224kPa 6:35  
136.4 C 224kPa 7:05  
136.4 C 222kPa 7:35  
136.4 C 222kPa 7:35  
VENTING 7:35  
AIR DRYING 8:30  
CYCLE COMPLETE 8:32

-----  
Digital Signature #  
3AEF074C1B36D389  
-----