

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2679

Name: RITTER 1  
Sterilizer ID: V2178198  
12/14/2022 04:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
2:00 204.1 F 4.3  
4:00 214.5 F 9.6  
6:00 225.0 F 6.4  
8:00 235.2 F 10.5  
10:00 234.9 F 7.4  
12:00 242.3 F 11.2  
14:00 246.6 F 13.4  
16:00 259.6 F 20.4  
18:00 271.1 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.0  
1:00 272.2 F 28.8  
2:00 271.7 F 28.3  
3:00 271.4 F 28.2  
4:00 272.3 F 28.8  
  
Min 271.1 F 27.9  
Max 272.4 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 233.7 F 0.0  
10:00 230.1 F 0.0  
15:00 227.6 F 0.0  
20:00 225.2 F 0.0  
25:00 223.9 F 0.0  
30:00 223.4 F 0.0  
  
Drying Complete

FILLING: 2:01  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:48  
CYCLE COMPLETE

12/14/2022 05:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_