
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001166
10:07 13/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:39
STERILIZING 3:46
136.0 C 220kPa 3:46
136.1 C 223kPa 4:16
136.4 C 224kPa 4:46
136.2 C 224kPa 5:16
136.3 C 224kPa 5:46
136.3 C 225kPa 6:16
136.3 C 225kPa 6:46
136.2 C 224kPa 7:16
136.2 C 224kPa 7:16
VENTING 7:16
AIR DRYING 8:08
CYCLE COMPLETE 23:08

Digital Signature #
94973B1D3E624C7E
