
Midmark M11 - v1.0.5
Total Cycles: 456

Name: RITTER2
Sterilizer ID: V2313228
10/18/2021 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.6 F 0.0
2:00 107.3 F 0.8
4:00 115.0 F 2.9
6:00 136.7 F 6.0
8:00 166.2 F 9.9
10:00 197.5 F 7.9
12:00 212.5 F 7.5
14:00 225.2 F 11.9
16:00 230.0 F 8.4
18:00 240.6 F 12.5
20:00 241.7 F 12.1
22:00 253.3 F 18.0
24:00 263.9 F 24.3
26:00 272.2 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.9
1:00 272.8 F 30.4
2:00 271.8 F 29.8
3:00 272.7 F 30.6
4:00 272.2 F 30.0

Min 271.7 F 29.8
Max 272.9 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 222.8 F 0.0
10:00 214.3 F 0.0
15:00 209.3 F 0.0
20:00 208.8 F 0.0
25:00 182.4 F 0.0
30:00 187.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:02
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:05:16
CYCLE COMPLETE

10/18/2021 10:45 AM
ACCEPTED: X
REJECTED:
Operator: U01